

DM-CAL

HEMATOLOGY CALIBRATOR

Assay values and ranges provided by DYMIND

CALIBRATOR

LOT N0122

2021-12-13

2022-02-10

Instrument	Parameter	Unit	Reference Value	Deviation
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31Vet, DH36Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.09	± 0.20
	RBC	$\times 10^{12}/L$	4.74	± 0.08
	HGB	g/L	139	± 2
	MCV	fL	95.3	± 2.0
	PLT	$\times 10^9/L$	246	± 12
DH20, DH21, DH22 DH23, DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.08	± 0.20
	RBC	$\times 10^{12}/L$	4.61	± 0.08
	HGB	g/L	137	± 2
	MCV	fL	87.0	± 2.0
	PLT	$\times 10^9/L$	249	± 12
DH31Vet, DH36Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.18	± 0.20
	RBC	$\times 10^{12}/L$	4.56	± 0.08
	HGB	g/L	137	± 2
	MCV	fL	96.6	± 2.0
	PLT	$\times 10^9/L$	245	± 12
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.31	± 0.20
	RBC	$\times 10^{12}/L$	4.60	± 0.08
	HGB	g/L	137	± 2
	MCV	fL	91.0	± 2.0
	PLT	$\times 10^9/L$	242	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park,Guangqiao road,Tianliao Community,Yutangstreet,Guangming District, Shenzhen, Guangdong,

P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132