



DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

 2022-01-10

 2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A9.0 to A11.9)	WBC	$\times 10^9/L$	9.37	± 0.20
	RBC	$\times 10^{12}/L$	4.42	± 0.08
	HGB	g/L	131	± 2
	MCV	fL	90.8	± 2.0
	PLT	$\times 10^9/L$	252	± 12
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11)	WBC	$\times 10^9/L$	9.93	± 0.20
	RBC	$\times 10^{12}/L$	4.42	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	86.6	± 2.0
	PLT	$\times 10^9/L$	249	± 12
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.12 or higher)	WBC	$\times 10^9/L$	10.27	± 0.20
	RBC	$\times 10^{12}/L$	4.46	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	86.2	± 2.0
	PLT	$\times 10^9/L$	239	± 12
UN71, UN73, UN76, DH73 (Technical File Version A2.0 to A5.3)	WBC	$\times 10^9/L$	9.64	± 0.20
	RBC	$\times 10^{12}/L$	4.47	± 0.08
	HGB	g/L	135	± 2
	MCV	fL	92.7	± 2.0
	PLT	$\times 10^9/L$	247	± 12

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park, Guangqiao road, Tianliao Community, Yutangstreet, Guangming District, Shenzhen, Guangdong, P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

2022-01-10

2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
UN71, UN73, UN76, DH73 (Technical File Version A5.5 to A6.2)	WBC	$\times 10^9/L$	9.47	± 0.20
	RBC	$\times 10^{12}/L$	4.37	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	88.3	± 2.0
	PLT	$\times 10^9/L$	245	± 12
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.30	± 0.20
	RBC	$\times 10^{12}/L$	4.50	± 0.08
	HGB	g/L	135	± 2
	MCV	fL	87.5	± 2.0
	PLT	$\times 10^9/L$	246	± 12
DH73Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.50	± 0.08
	HGB	g/L	135	± 2
	MCV	fL	92.5	± 2.0
	PLT	$\times 10^9/L$	253	± 12
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A10.0 to A11.1) DF50 Vet, DF52 Vet (Technical File Version A2.0 to A7.0)	WBC	$\times 10^9/L$	9.07	± 0.20
	RBC	$\times 10^{12}/L$	4.55	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	88.7	± 2.0
	PLT	$\times 10^9/L$	256	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

2022-01-10

2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A11.2 to A11.6)	WBC	$\times 10^9/L$	9.07	± 0.20
	RBC	$\times 10^{12}/L$	4.56	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	89.7	± 2.0
	PLT	$\times 10^9/L$	253	± 12
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	9.15	± 0.20
	RBC	$\times 10^{12}/L$	4.59	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	86.9	± 2.0
	PLT	$\times 10^9/L$	246	± 12
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.18	± 0.20
	RBC	$\times 10^{12}/L$	4.58	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	86.9	± 2.0
	PLT	$\times 10^9/L$	255	± 12
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.29	± 0.20
	RBC	$\times 10^{12}/L$	4.53	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	93.0	± 2.0
	PLT	$\times 10^9/L$	257	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park, Guangqiao road, Tianliao Community, Yutangstreet, Guangming District, Shenzhen, Guangdong, P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

2022-01-10

2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
DH71, DH76 (Technical File Version A10.4)	WBC	$\times 10^9/L$	8.67	± 0.20
	RBC	$\times 10^{12}/L$	4.61	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	88.4	± 2.0
	PLT	$\times 10^9/L$	254	± 12
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	9.56	± 0.20
	RBC	$\times 10^{12}/L$	4.54	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	86.5	± 2.0
	PLT	$\times 10^9/L$	250	± 12
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A3.0 to A5.11)	WBC	$\times 10^9/L$	9.23	± 0.20
	RBC	$\times 10^{12}/L$	4.46	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	91.7	± 2.0
	PLT	$\times 10^9/L$	244	± 12
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.28	± 0.20
	RBC	$\times 10^{12}/L$	4.41	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	88.1	± 2.0
	PLT	$\times 10^9/L$	254	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

2022-01-10

2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	9.59	± 0.20
	RBC	$\times 10^{12}/L$	4.48	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	86.8	± 2.0
	PLT	$\times 10^9/L$	246	± 12
D6-CRP, DH76CRP (Technical File Version B1.0 to B2.0)	WBC	$\times 10^9/L$	9.17	± 0.20
	RBC	$\times 10^{12}/L$	4.40	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	87.9	± 2.0
	PLT	$\times 10^9/L$	248	± 12
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	9.19	± 0.20
	RBC	$\times 10^{12}/L$	4.51	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	86.9	± 2.0
	PLT	$\times 10^9/L$	257	± 12
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	9.13	± 0.20
	RBC	$\times 10^{12}/L$	4.59	± 0.08
	HGB	g/L	132	± 2
	MCV	fL	90.6	± 2.0
	PLT	$\times 10^9/L$	253	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park, Guangqiao road, Tianliao Community, Yutangstreet, Guangming District, Shenzhen, Guangdong, P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT NK0222

2022-01-10

2022-03-10

Instrument	Parameter	Unit	Reference Value	Deviation
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A4.0 to A5.7)	WBC	$\times 10^9/L$	9.10	± 0.20
	RBC	$\times 10^{12}/L$	4.48	± 0.08
	HGB	g/L	131	± 2
	MCV	fL	90.1	± 2.0
	PLT	$\times 10^9/L$	256	± 12
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.0)	WBC	$\times 10^9/L$	9.58	± 0.20
	RBC	$\times 10^{12}/L$	4.46	± 0.08
	HGB	g/L	131	± 2
	MCV	fL	87.6	± 2.0
	PLT	$\times 10^9/L$	256	± 12
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	9.65	± 0.20
	RBC	$\times 10^{12}/L$	4.45	± 0.08
	HGB	g/L	133	± 2
	MCV	fL	87.5	± 2.0
	PLT	$\times 10^9/L$	251	± 12
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.18	± 0.20
	RBC	$\times 10^{12}/L$	4.52	± 0.08
	HGB	g/L	134	± 2
	MCV	fL	87.6	± 2.0
	PLT	$\times 10^9/L$	252	± 12

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature (15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

Shenzhen Dymind Biotechnology Co., Ltd

Add: B of 12/F, High-tech park, Guangqiao road, Tianliao Community, Yutangstreet, Guangming District, Shenzhen, Guangdong, P.R.China.

Tel: +86 755 26008015

Fax: +86 755 26746162

Postcode: 518132