

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (Technical File Version A9.0 to A11.9)	WBC	×10 ⁹ /L	3.35	±0.50	8.04	±1.00	17.89	±2.50
	Neu%	%	51.3	±9.0	63.9	±8.0	76.0	±23.0
	Lym%	%	37.8	±9.0	26.0	±8.0	15.0	±15.0
	Mon%	%	6.9	±4.0	6.1	±5.0	5.0	±5.0
	Eos%	%	4.0	±4.0	4.0	±4.0	4.0	±4.0
	Bas%	%	63.6	±8.0	72.0	±8.0	80.5	±8.0
	Neu#	×10 ⁹ /L	1.72	±0.40	5.14	±0.70	13.60	±6.00
	Lym#	×10 ⁹ /L	1.27	±0.40	2.09	±0.70	2.68	±2.68
	Mon#	×10 ⁹ /L	0.23	±0.14	0.49	±0.49	0.89	±0.89
	Eos#	×10 ⁹ /L	0.13	±0.13	0.32	±0.32	0.72	±0.72
	Bas#	×10 ⁹ /L	2.13	±0.30	5.79	±0.70	14.40	±1.50
	RBC	×10 ¹² /L	2.12	±0.18	4.45	±0.24	5.31	±0.50
	HGB	g/L	57	±4	129	±6	167	±8
	HCT	%	18.2	±2.0	40.8	±3.0	53.3	±4.0
	MCV	fL	85.7	±5.0	91.7	±5.0	100.4	±6.0
	MCH	pg	26.5	±2.5	28.6	±2.5	31.0	±2.5
	MCHC	g/L	310	±30	313	±30	309	±30
	RDW-CV	%	18.5	±3.0	17.2	±3.0	16.4	±3.0
	RDW-SD	fL	55.8	±10.0	54.7	±10.0	56.5	±12.0
	PLT	×10 ⁹ /L	50	±20	253	±40	526	±60
	MPV	fL	10.1	±3.0	10.5	±3.0	9.4	±3.0
	PDW	fL	10.6	±3.0	13.6	±3.0	11.2	±3.0
	PCT	%	0.051	±0.050	0.271	±0.100	0.499	±0.200
P-LCR	%	24.5	±8.0	30.5	±8.0	22.7	±8.0	
P-LCC	×10 ⁹ /L	15	±15	79	±25	121	±35	
PDW	/	10.2	±3.0	10.3	±3.0	10.2	±3.0	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (Technical File Version A11.11)	WBC	×10 ⁹ /L	3.45	±0.50	8.52	±1.00	18.91	±2.50
	Neu%	%	51.5	±9.0	61.1	±8.0	73.5	±20.0
	Lym%	%	38.4	±9.0	28.9	±8.0	16.9	±10.0
	Mon%	%	6.1	±4.0	6.0	±5.0	5.6	±5.6
	Eos%	%	4.0	±4.0	4.0	±4.0	4.0	±4.0
	Bas%	%	64.2	±8.0	72.2	±8.0	81.8	±8.0
	Neu#	×10 ⁹ /L	1.78	±0.40	5.20	±0.70	13.90	±4.00
	Lym#	×10 ⁹ /L	1.32	±0.40	2.46	±0.70	3.20	±2.00
	Mon#	×10 ⁹ /L	0.21	±0.14	0.51	±0.50	1.06	±1.06
	Eos#	×10 ⁹ /L	0.14	±0.14	0.34	±0.34	0.76	±0.76
	Bas#	×10 ⁹ /L	2.21	±0.30	6.15	±0.70	15.47	±1.50
	RBC	×10 ¹² /L	2.16	±0.18	4.47	±0.24	5.37	±0.50
	HGB	g/L	56	±4	128	±6	165	±8
	HCT	%	17.7	±2.0	39.0	±3.0	50.8	±4.0
	MCV	fL	82.2	±5.0	87.2	±5.0	94.6	±6.0
	MCH	pg	26.4	±2.5	28.8	±2.5	31.0	±2.5
	MCHC	g/L	322	±30	331	±30	328	±30
	RDW-CV	%	18.2	±3.0	16.8	±3.0	15.8	±3.0
	RDW-SD	fL	54.7	±10.0	53.6	±10.0	54.2	±12.0
	PLT	×10 ⁹ /L	48	±20	245	±40	518	±60
	MPV	fL	10.3	±3.0	10.6	±3.0	9.5	±3.0
PDW	fL	11.2	±3.0	13.6	±3.0	11.4	±3.0	
PCT	%	0.050	±0.050	0.263	±0.100	0.503	±0.200	
P-LCR	%	25.7	±8.0	30.9	±8.0	23.5	±8.0	
P-LCC	×10 ⁹ /L	15	±15	77	±25	124	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (Technical File Version A11.12 or higher)	WBC	×10 ⁹ /L	3.53	±0.50	8.83	±1.00	19.50	±2.50
	Neu%	%	51.6	±9.0	60.0	±8.0	68.8	±10.0
	Lym%	%	38.5	±9.0	29.3	±8.0	19.4	±6.0
	Mon%	%	6.9	±4.0	6.7	±5.0	6.0	±6.0
	Eos%	%	3.0	±3.0	4.0	±4.0	5.8	±5.8
	Bas%	%	64.3	±8.0	72.2	±8.0	80.2	±8.0
	Neu#	×10 ⁹ /L	1.82	±0.40	5.30	±0.70	13.42	±2.00
	Lym#	×10 ⁹ /L	1.36	±0.40	2.59	±0.70	3.78	±1.10
	Mon#	×10 ⁹ /L	0.24	±0.14	0.59	±0.50	1.17	±1.10
	Eos#	×10 ⁹ /L	0.11	±0.11	0.35	±0.35	1.13	±1.13
	Bas#	×10 ⁹ /L	2.27	±0.30	6.38	±0.70	15.64	±1.50
	RBC	×10 ¹² /L	2.16	±0.18	4.48	±0.24	5.39	±0.50
	HGB	g/L	57	±4	129	±6	167	±8
	HCT	%	17.8	±2.0	39.1	±3.0	51.2	±4.0
	MCV	fL	82.4	±5.0	87.3	±5.0	95.0	±6.0
	MCH	pg	26.5	±2.5	28.9	±2.5	31.2	±2.5
	MCHC	g/L	322	±30	331	±30	329	±30
	RDW-CV	%	18.1	±3.0	16.8	±3.0	15.8	±3.0
	RDW-SD	fL	54.3	±10.0	53.3	±10.0	54.0	±12.0
	PLT	×10 ⁹ /L	49	±20	250	±40	517	±60
MPV	fL	10.0	±3.0	10.4	±3.0	9.2	±3.0	
PDW	fL	10.5	±3.0	13.4	±3.0	10.8	±3.0	
PCT	%	0.051	±0.050	0.265	±0.100	0.495	±0.200	
P-LCR	%	23.9	±8.0	29.7	±8.0	21.8	±8.0	
P-LCC	×10 ⁹ /L	15	±15	76	±25	117	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND UN71 UN73 UN76 DH73 (Technical File Version A2.0 to A5.3)	WBC	×10 ⁹ /L	3.50	±0.50	8.40	±1.00	18.40	±2.50
	Neu%	%	50.4	±9.0	57.8	±8.0	66.2	±7.0
	Lym%	%	37.9	±9.0	29.1	±8.0	19.5	±6.0
	Mon%	%	6.7	±4.0	7.1	±5.0	7.2	±6.0
	Eos%	%	5.0	±5.0	6.0	±6.0	7.1	±7.0
	Bas%	%	0.2	±0.2	0.5	±0.5	0.2	±0.2
	Neu#	×10 ⁹ /L	1.76	±0.40	4.86	±0.70	12.18	±1.40
	Lym#	×10 ⁹ /L	1.33	±0.40	2.44	±0.70	3.59	±1.10
	Mon#	×10 ⁹ /L	0.23	±0.14	0.60	±0.50	1.32	±1.10
	Eos#	×10 ⁹ /L	0.18	±0.15	0.50	±0.50	1.31	±1.30
	Bas#	×10 ⁹ /L	0.01	±0.01	0.04	±0.04	0.04	±0.04
	RBC	×10 ¹² /L	2.17	±0.18	4.51	±0.24	5.37	±0.50
	HGB	g/L	59	±4	131	±6	171	±8
	HCT	%	18.9	±2.0	41.5	±3.0	54.5	±4.0
	MCV	fL	86.8	±5.0	92.1	±5.0	101.4	±6.0
	MCH	pg	27.0	±2.5	28.8	±2.5	31.5	±2.5
	MCHC	g/L	314	±30	313	±30	312	±30
	RDW-CV	%	18.6	±3.0	17.3	±3.0	16.5	±3.0
	RDW-SD	fL	56.2	±10.0	55.3	±10.0	57.1	±12.0
	PLT	×10 ⁹ /L	57	±20	249	±40	502	±60
MPV	fL	9.7	±3.0	10.7	±3.0	9.6	±3.0	
PDW	fL	10.3	±3.0	10.4	±3.0	10.3	±3.0	
PCT	%	0.056	±0.050	0.270	±0.100	0.485	±0.200	
P-LCR	%	26.1	±8.0	32.4	±8.0	24.1	±8.0	
P-LCC	×10 ⁹ /L	15	±15	82	±25	122	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND UN71 UN73 UN76 DH73 Technical File Version A5.5 to A6.2)	WBC	×10 ⁹ /L	3.51	±0.50	8.25	±1.00	18.02	±2.50
	Neu%	%	51.4	±9.0	56.8	±8.0	66.6	±7.0
	Lym%	%	39.4	±9.0	31.1	±8.0	21.0	±6.0
	Mon%	%	6.7	±4.0	7.1	±5.0	7.4	±6.0
	Eos%	%	1.5	±1.5	4.0	±4.0	4.0	±4.0
	Bas%	%	1.0	±1.0	1.0	±1.0	1.0	±1.0
	Neu#	×10 ⁹ /L	1.80	±0.40	4.69	±0.70	12.00	±1.40
	Lym#	×10 ⁹ /L	1.38	±0.40	2.57	±0.70	3.79	±1.10
	Mon#	×10 ⁹ /L	0.24	±0.14	0.59	±0.50	1.33	±1.10
	Eos#	×10 ⁹ /L	0.05	±0.05	0.33	±0.33	0.72	±0.72
	Bas#	×10 ⁹ /L	0.04	±0.04	0.08	±0.08	0.18	±0.18
	RBC	×10 ¹² /L	2.14	±0.18	4.40	±0.24	5.19	±0.50
	HGB	g/L	59	±4	130	±6	168	±8
	HCT	%	17.5	±2.0	38.7	±3.0	49.4	±4.0
	MCV	fL	82.0	±5.0	88.0	±5.0	95.1	±6.0
	MCH	pg	27.4	±2.5	29.5	±2.5	32.0	±2.5
	MCHC	g/L	331	±30	333	±30	334	±30
	RDW-CV	%	19.0	±3.0	17.7	±3.0	17.1	±3.0
	RDW-SD	fL	58.9	±10.0	58.1	±10.0	60.5	±12.0
	PLT	×10 ⁹ /L	55	±20	246	±40	489	±60
MPV	fL	9.3	±3.0	10.4	±3.0	9.3	±3.0	
PDW	fL	12.0	±3.0	13.2	±3.0	11.0	±3.0	
PCT	%	0.051	±0.050	0.255	±0.100	0.453	±0.200	
P-LCR	%	23.3	±8.0	29.6	±8.0	22.0	±8.0	
P-LCC	×10 ⁹ /L	15	±15	73	±25	108	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low	Normal	High
DYMIND UN71 UN73 UN76 DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	3.42 ± 0.50	8.17 ± 1.00	17.95 ± 2.50
	Neu%	%	48.3 ± 9.0	56.9 ± 8.0	66.1 ± 7.0
	Lym%	%	39.0 ± 9.0	31.2 ± 8.0	20.6 ± 6.0
	Mon%	%	6.7 ± 4.0	6.9 ± 5.0	7.3 ± 6.0
	Eos%	%	5.0 ± 5.0	4.0 ± 4.0	5.0 ± 5.0
	Bas%	%	1.0 ± 1.0	1.0 ± 1.0	1.0 ± 1.0
	Neu#	$\times 10^9/L$	1.65 ± 0.40	4.65 ± 0.70	11.86 ± 1.40
	Lym#	$\times 10^9/L$	1.33 ± 0.40	2.55 ± 0.70	3.70 ± 1.10
	Mon#	$\times 10^9/L$	0.23 ± 0.14	0.56 ± 0.50	1.31 ± 1.10
	Eos#	$\times 10^9/L$	0.17 ± 0.15	0.33 ± 0.33	0.90 ± 0.90
	Bas#	$\times 10^9/L$	0.03 ± 0.03	0.08 ± 0.08	0.18 ± 0.18
	RBC	$\times 10^{12}/L$	2.24 ± 0.18	4.53 ± 0.24	5.44 ± 0.50
	HGB	g/L	59 ± 4	131 ± 6	170 ± 8
	HCT	%	18.4 ± 2.0	39.9 ± 3.0	52.1 ± 4.0
	MCV	fL	82.0 ± 5.0	88.1 ± 5.0	95.9 ± 6.0
	MCH	pg	26.6 ± 2.5	29.0 ± 2.5	31.4 ± 2.5
	MCHC	g/L	323 ± 30	327 ± 30	326 ± 30
	RDW-CV	%	18.1 ± 3.0	16.8 ± 3.0	16.0 ± 3.0
	RDW-SD	fL	56.1 ± 10.0	55.0 ± 10.0	56.3 ± 12.0
	PLT	$\times 10^9/L$	56 ± 20	245 ± 40	492 ± 60
MPV	fL	9.6 ± 3.0	10.4 ± 3.0	9.3 ± 3.0	
PDW	fL	12.3 ± 3.0	13.2 ± 3.0	11.0 ± 3.0	
PCT	%	0.053 ± 0.050	0.253 ± 0.100	0.448 ± 0.200	
P-LCR	%	25.5 ± 8.0	29.9 ± 8.0	22.0 ± 8.0	
P-LCC	$\times 10^9/L$	15 ± 15	72 ± 25	107 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	2.20	± 0.18	4.52	± 0.24	5.38	± 0.50
	HGB	g/L	59	± 4	131	± 6	171	± 8
	HCT	%	19.0	± 2.0	41.6	± 3.0	54.3	± 4.0
	MCV	fL	86.3	± 5.0	92.1	± 5.0	100.9	± 6.0
	MCH	pg	26.9	± 2.5	29.1	± 2.5	31.8	± 2.5
	MCHC	g/L	312	± 30	316	± 30	315	± 30
	RDW-CV	%	18.1	± 3.0	16.8	± 3.0	15.9	± 3.0
	RDW-SD	fL	54.8	± 10.0	53.6	± 10.0	54.9	± 12.0
	PLT	$\times 10^9/L$	57	± 20	245	± 40	489	± 60
	MPV	fL	9.7	± 3.0	10.5	± 3.0	9.3	± 3.0
	PDW	fL	12.3	± 3.0	13.4	± 3.0	11.2	± 3.0
	PCT	%	0.056	± 0.050	0.257	± 0.100	0.456	± 0.200
	P-LCR	%	26.0	± 8.0	30.3	± 8.0	22.7	± 8.0
	P-LCC	$\times 10^9/L$	15	± 15	74	± 25	111	± 35

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version A10.0 to A11.1) DF50 Vet DF52 Vet (Technical File Version A2.0 to A7.0)	WBC	×10 ⁹ /L	3.41	±0.50	8.05	±1.00	18.04	±2.50
	Neu%	%	49.7	±9.0	56.3	±8.0	62.4	±7.0
	Lym%	%	39.1	±9.0	30.3	±8.0	20.2	±6.0
	Mon%	%	5.2	±4.0	5.0	±5.0	5.5	±5.5
	Eos%	%	6.0	±5.0	8.4	±6.0	11.9	±7.0
	Bas%	%	2.5	±2.5	1.0	±1.0	2.0	±2.0
	Neu#	×10 ⁹ /L	1.69	±0.40	4.53	±0.70	11.25	±1.40
	Lym#	×10 ⁹ /L	1.33	±0.40	2.44	±0.70	3.64	±1.10
	Mon#	×10 ⁹ /L	0.18	±0.14	0.40	±0.40	0.99	±0.99
	Eos#	×10 ⁹ /L	0.20	±0.15	0.68	±0.50	2.15	±1.30
	Bas#	×10 ⁹ /L	0.09	±0.09	0.08	±0.08	0.36	±0.36
	RBC	×10 ¹² /L	2.18	±0.18	4.55	±0.24	5.39	±0.50
	HGB	g/L	59	±4	130	±6	172	±8
	HCT	%	17.5	±2.0	39.8	±3.0	51.6	±4.0
	MCV	fL	81.6	±5.0	88.2	±5.0	96.1	±6.0
	MCH	pg	26.7	±2.5	28.4	±2.5	31.6	±2.5
	MCHC	g/L	332	±30	323	±30	330	±30
	RDW-CV	%	15.5	±3.0	15.0	±3.0	14.6	±3.0
	RDW-SD	fL	52.2	±10.0	53.4	±10.0	56.3	±12.0
	PLT	×10 ⁹ /L	68	±20	257	±40	493	±60
MPV	fL	9.4	±3.0	10.1	±3.0	9.2	±3.0	
PDW	/	15.0	±3.0	16.2	±3.0	16.3	±3.0	
PCT	%	0.064	±0.050	0.262	±0.100	0.459	±0.200	
P-LCR	%	29.9	±8.0	38.6	±8.0	31.2	±8.0	
P-LCC	×10 ⁹ /L	21	±15	100	±25	155	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not use on DF50,DF51,DF53, DF50 Vet and DF52 Vet model.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version A11.2 to A11.6)	WBC	×10 ⁹ /L	3.42	±0.50	8.15	±1.00	17.99	±2.50
	Neu%	%	49.7	±9.0	55.5	±8.0	62.5	±7.0
	Lym%	%	39.3	±9.0	30.8	±8.0	20.3	±6.0
	Mon%	%	5.2	±4.0	5.0	±5.0	5.2	±5.2
	Eos%	%	5.8	±5.0	8.7	±6.0	12.0	±7.0
	Bas%	%	2.5	±2.5	1.5	±1.5	2.0	±2.0
	Neu#	×10 ⁹ /L	1.70	±0.40	4.52	±0.70	11.25	±1.40
	Lym#	×10 ⁹ /L	1.34	±0.40	2.51	±0.70	3.65	±1.10
	Mon#	×10 ⁹ /L	0.18	±0.14	0.41	±0.41	0.94	±0.94
	Eos#	×10 ⁹ /L	0.20	±0.15	0.71	±0.50	2.16	±1.30
	Bas#	×10 ⁹ /L	0.09	±0.09	0.12	±0.12	0.36	±0.36
	RBC	×10 ¹² /L	2.23	±0.18	4.61	±0.24	5.47	±0.50
	HGB	g/L	59	±4	132	±6	171	±8
	HCT	%	18.6	±2.0	41.3	±3.0	52.9	±4.0
	MCV	fL	83.4	±5.0	89.6	±5.0	96.7	±6.0
	MCH	pg	26.3	±2.5	28.6	±2.5	31.2	±2.5
	MCHC	g/L	317	±30	320	±30	324	±30
	RDW-CV	%	15.0	±3.0	14.1	±3.0	13.4	±3.0
	RDW-SD	fL	52.2	±10.0	52.0	±10.0	52.9	±12.0
	PLT	×10 ⁹ /L	60	±20	254	±40	495	±60
MPV	fL	9.9	±3.0	10.6	±3.0	9.5	±3.0	
PDW	fL	12.5	±3.0	15.4	±3.0	13.2	±3.0	
PCT	%	0.059	±0.050	0.272	±0.100	0.472	±0.200	
P-LCR	%	32.9	±8.0	41.7	±8.0	33.7	±8.0	
P-LCC	×10 ⁹ /L	20	±15	107	±25	167	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not use on DF50,DF51 and DF53 model.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low	Normal	High
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	3.45 ± 0.50	8.12 ± 1.00	18.23 ± 2.50
	Neu%	%	49.9 ± 9.0	56.1 ± 8.0	62.3 ± 7.0
	Lym%	%	39.3 ± 9.0	30.4 ± 8.0	20.2 ± 6.0
	Mon%	%	5.1 ± 4.0	5.1 ± 5.0	5.7 ± 5.7
	Eos%	%	5.7 ± 5.0	8.4 ± 6.0	11.8 ± 7.0
	Bas%	%	2.5 ± 2.5	1.0 ± 1.0	2.0 ± 2.0
	Neu#	$\times 10^9/L$	1.72 ± 0.40	4.56 ± 0.70	11.36 ± 1.40
	Lym#	$\times 10^9/L$	1.36 ± 0.40	2.47 ± 0.70	3.68 ± 1.10
	Mon#	$\times 10^9/L$	0.18 ± 0.14	0.41 ± 0.41	1.04 ± 1.04
	Eos#	$\times 10^9/L$	0.20 ± 0.15	0.68 ± 0.50	2.15 ± 1.30
	Bas#	$\times 10^9/L$	0.09 ± 0.09	0.08 ± 0.08	0.36 ± 0.36
	RBC	$\times 10^{12}/L$	2.23 ± 0.18	4.58 ± 0.24	5.47 ± 0.50
	HGB	g/L	59 ± 4	130 ± 6	173 ± 8
	HCT	%	18.7 ± 2.0	40.0 ± 3.0	51.3 ± 4.0
	MCV	fL	83.9 ± 5.0	87.4 ± 5.0	93.9 ± 6.0
	MCH	pg	26.4 ± 2.5	28.6 ± 2.5	31.6 ± 2.5
	MCHC	g/L	317 ± 30	327 ± 30	337 ± 30
	RDW-CV	%	15.6 ± 3.0	14.8 ± 3.0	14.1 ± 3.0
	RDW-SD	fL	55.6 ± 10.0	55.7 ± 10.0	57.3 ± 12.0
	PLT	$\times 10^9/L$	59 ± 20	247 ± 40	475 ± 60
MPV	fL	9.3 ± 3.0	10.0 ± 3.0	9.2 ± 3.0	
PDW	fL	11.8 ± 3.0	14.6 ± 3.0	12.3 ± 3.0	
PCT	%	0.055 ± 0.050	0.248 ± 0.100	0.434 ± 0.200	
P-LCR	%	29.4 ± 8.0	38.2 ± 8.0	30.5 ± 8.0	
P-LCC	$\times 10^9/L$	17 ± 15	94 ± 25	145 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not use on DF50,DF51 and DF53 model.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low	Normal	High
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	3.39 ± 0.50	8.01 ± 1.00	18.01 ± 2.50
	Neu%	%	50.5 ± 9.0	56.3 ± 8.0	62.6 ± 7.0
	Lym%	%	38.5 ± 9.0	29.8 ± 8.0	19.9 ± 6.0
	Mon%	%	5.4 ± 4.0	5.2 ± 5.0	5.5 ± 5.5
	Eos%	%	5.6 ± 5.0	8.7 ± 6.0	12.0 ± 7.0
	Bas%	%	2.5 ± 2.5	1.0 ± 1.0	2.0 ± 2.0
	Neu#	$\times 10^9/L$	1.71 ± 0.40	4.51 ± 0.70	11.28 ± 1.40
	Lym#	$\times 10^9/L$	1.31 ± 0.40	2.39 ± 0.70	3.58 ± 1.10
	Mon#	$\times 10^9/L$	0.18 ± 0.14	0.42 ± 0.42	0.99 ± 0.99
	Eos#	$\times 10^9/L$	0.19 ± 0.15	0.70 ± 0.50	2.16 ± 1.30
	Bas#	$\times 10^9/L$	0.08 ± 0.08	0.08 ± 0.08	0.36 ± 0.36
	RBC	$\times 10^{12}/L$	2.22 ± 0.18	4.57 ± 0.24	5.43 ± 0.50
	HGB	g/L	58 ± 4	132 ± 6	172 ± 8
	HCT	%	18.2 ± 2.0	39.6 ± 3.0	50.5 ± 4.0
	MCV	fL	82.0 ± 5.0	86.6 ± 5.0	93.1 ± 6.0
	MCH	pg	27 ± 2.5	29 ± 2.5	32 ± 2.5
	MCHC	g/L	324 ± 30	334 ± 30	344 ± 30
	RDW-CV	%	15.3 ± 3.0	14.4 ± 3.0	13.7 ± 3.0
	RDW-SD	fL	53.0 ± 10.0	52.8 ± 10.0	53.8 ± 12.0
	PLT	$\times 10^9/L$	63 ± 20	264 ± 40	493 ± 60
	MPV	fL	9.1 ± 3.0	9.7 ± 3.0	8.8 ± 3.0
PDW	fL	11.5 ± 3.0	15.4 ± 3.0	12.8 ± 3.0	
PCT	%	0.058 ± 0.050	0.262 ± 0.100	0.439 ± 0.200	
P-LCR	%	26.4 ± 8.0	35.9 ± 8.0	28.2 ± 8.0	
P-LCC	$\times 10^9/L$	17 ± 15	96 ± 25	141 ± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.
5. P-LCC and P-LCR parameters are not use on DF50,DF51 and DF53 model.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND	WBC	$\times 10^9/L$	3.41	± 0.50	7.99	± 1.00	17.64	± 2.50
	Neu%	%	50.8	± 9.0	57.9	± 8.0	67.7	± 7.0
DH71	Lym%	%	38.2	± 9.0	29.7	± 8.0	19.7	± 6.0
DH76	Mon%	%	7.0	± 4.0	7.4	± 5.0	7.6	± 6.0
(Technical File Version A4.1 to A10.3)	Eos%	%	4.0	± 4.0	5.0	± 5.0	5.0	± 5.0
	Bas%	%	63.1	± 8.0	71.2	± 8.0	80.9	± 8.0
DH51	Neu#	$\times 10^9/L$	1.73	± 0.40	4.62	± 0.70	11.94	± 1.40
DH53	Lym#	$\times 10^9/L$	1.30	± 0.40	2.37	± 0.70	3.48	± 1.10
DH56	Mon#	$\times 10^9/L$	0.24	± 0.14	0.59	± 0.50	1.34	± 1.10
(Technical File Version A8.1 or higher)	Eos#	$\times 10^9/L$	0.14	± 0.14	0.40	± 0.40	0.88	± 0.88
	Bas#	$\times 10^9/L$	2.15	± 0.30	5.69	± 0.70	14.27	± 1.50
	RBC	$\times 10^{12}/L$	2.21	± 0.18	4.57	± 0.24	5.51	± 0.50
	HGB	g/L	58	± 4	130	± 6	168	± 8
	HCT	%	19.3	± 2.0	42.5	± 3.0	56.2	± 4.0
	MCV	fL	87.0	± 5.0	92.9	± 5.0	101.9	± 6.0
	MCH	pg	26.0	± 2.5	28.6	± 2.5	30.5	± 2.5
	MCHC	g/L	297	± 30	306	± 30	298	± 30
	RDW-CV	%	18.1	± 3.0	16.8	± 3.0	15.7	± 3.0
	RDW-SD	fL	55.8	± 10.0	54.8	± 10.0	55.5	± 12.0
	PLT	$\times 10^9/L$	56	± 20	252	± 40	533	± 60
	MPV	fL	9.8	± 3.0	10.7	± 3.0	9.6	± 3.0
	PDW	/	10.3	± 3.0	10.3	± 3.0	10.2	± 3.0
	PCT	%	0.056	± 0.050	0.275	± 0.100	0.520	± 0.200
	P-LCR	%	26.7	± 8.0	32.1	± 8.0	24.0	± 8.0
	P-LCC	$\times 10^9/L$	15	± 15	82	± 25	131	± 35

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DH71 DH76 (Technical File Version A10.4)	WBC	×10 ⁹ /L	3.26	±0.50	7.47	±1.00	16.41	±2.50
	Neu%	%	50.0	±9.0	56.8	±8.0	66.5	±7.0
	Lym%	%	38.9	±9.0	31.0	±8.0	20.9	±6.0
	Mon%	%	7.1	±4.0	7.2	±5.0	7.6	±6.0
	Eos%	%	4.0	±4.0	5.0	±5.0	5.0	±5.0
	Bas%	%	62.1	±8.0	70.2	±8.0	80.3	±8.0
	Neu#	×10 ⁹ /L	1.63	±0.40	4.24	±0.70	10.91	±1.40
	Lym#	×10 ⁹ /L	1.27	±0.40	2.32	±0.70	3.43	±1.10
	Mon#	×10 ⁹ /L	0.23	±0.14	0.54	±0.50	1.25	±1.10
	Eos#	×10 ⁹ /L	0.13	±0.13	0.37	±0.37	0.82	±0.82
	Bas#	×10 ⁹ /L	2.02	±0.30	5.24	±0.70	13.18	±1.50
	RBC	×10 ¹² /L	2.21	±0.18	4.56	±0.24	5.51	±0.50
	HGB	g/L	57	±4	130	±6	168	±8
	HCT	%	18.3	±2.0	40.2	±3.0	52.7	±4.0
	MCV	fL	82.7	±5.0	88.2	±5.0	95.7	±6.0
	MCH	pg	25.9	±2.5	28.6	±2.5	30.5	±2.5
	MCHC	g/L	313	±30	324	±30	319	±30
	RDW-CV	%	18.2	±3.0	16.9	±3.0	15.9	±3.0
	RDW-SD	fL	55.2	±10.0	54.3	±10.0	55.1	±12.0
	PLT	×10 ⁹ /L	47	±20	248	±40	522	±60
MPV	fL	10.4	±3.0	10.6	±3.0	9.5	±3.0	
PDW	fL	11.2	±3.0	13.7	±3.0	11.5	±3.0	
PCT	%	0.049	±0.049	0.264	±0.100	0.502	±0.200	
P-LCR	%	26.2	±8.0	31.1	±8.0	23.4	±8.0	
P-LCC	×10 ⁹ /L	15	±15	78	±25	124	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low	Normal	High
DYMIND DH71 DH76 (Technical File Version A10.5 or higher)	WBC	×10 ⁹ /L	3.34 ±0.50	8.00 ±1.00	18.03 ±2.50
	Neu%	%	50.2 ±9.0	57.8 ±8.0	66.9 ±7.0
	Lym%	%	38.7 ±9.0	31.0 ±8.0	20.4 ±6.0
	Mon%	%	7.1 ±4.0	7.2 ±5.0	7.7 ±6.0
	Eos%	%	4.0 ±4.0	4.0 ±4.0	5.0 ±5.0
	Bas%	%	65.0 ±8.0	72.8 ±8.0	82.8 ±8.0
	Neu#	×10 ⁹ /L	1.68 ±0.40	4.63 ±0.70	12.06 ±1.40
	Lym#	×10 ⁹ /L	1.29 ±0.40	2.48 ±0.70	3.68 ±1.10
	Mon#	×10 ⁹ /L	0.24 ±0.14	0.58 ±0.50	1.39 ±1.10
	Eos#	×10 ⁹ /L	0.13 ±0.13	0.32 ±0.32	0.90 ±0.90
	Bas#	×10 ⁹ /L	2.17 ±0.30	5.83 ±0.70	14.93 ±1.50
	RBC	×10 ¹² /L	2.19 ±0.18	4.52 ±0.24	5.48 ±0.50
	HGB	g/L	58 ±4	130 ±6	169 ±8
	HCT	%	17.9 ±2.0	39.5 ±3.0	52.3 ±4.0
	MCV	fL	81.7 ±5.0	87.3 ±5.0	95.5 ±6.0
	MCH	pg	26.6 ±2.5	29.1 ±2.5	31.3 ±2.5
	MCHC	g/L	324 ±30	332 ±30	327 ±30
	RDW-CV	%	18.2 ±3.0	16.8 ±3.0	15.7 ±3.0
	RDW-SD	fL	54.8 ±10.0	53.7 ±10.0	54.1 ±12.0
	PLT	×10 ⁹ /L	46 ±20	243 ±40	508 ±60
MPV	fL	10.0 ±3.0	10.3 ±3.0	9.1 ±3.0	
PDW	fL	10.9 ±3.0	13.1 ±3.0	10.8 ±3.0	
PCT	%	0.047 ±0.047	0.250 ±0.100	0.462 ±0.200	
P-LCR	%	24.3 ±8.0	29.2 ±8.0	21.1 ±8.0	
P-LCC	×10 ⁹ /L	15 ±15	71 ±25	107 ±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (Technical File Version A3.0 to A5.11)	WBC	×10 ⁹ /L	3.38	±0.50	7.94	±1.00	17.54	±2.50
	Neu%	%	50.8	±9.0	58.5	±8.0	68.6	±7.0
	Lym%	%	38.3	±9.0	29.6	±8.0	19.4	±6.0
	Mon%	%	6.9	±4.0	6.9	±5.0	7.0	±6.0
	Eos%	%	4.0	±4.0	5.0	±5.0	5.0	±5.0
	Bas%	%	64.1	±8.0	71.5	±8.0	81.5	±8.0
	Neu#	×10 ⁹ /L	1.72	±0.40	4.64	±0.70	12.04	±1.40
	Lym#	×10 ⁹ /L	1.29	±0.40	2.35	±0.70	3.40	±1.10
	Mon#	×10 ⁹ /L	0.23	±0.14	0.55	±0.50	1.23	±1.10
	Eos#	×10 ⁹ /L	0.14	±0.14	0.40	±0.40	0.88	±0.88
	Bas#	×10 ⁹ /L	2.17	±0.30	5.68	±0.70	14.30	±1.50
	RBC	×10 ¹² /L	2.17	±0.18	4.44	±0.24	5.28	±0.50
	HGB	g/L	57	±4	128	±6	165	±8
	HCT	%	18.6	±2.0	40.6	±3.0	52.9	±4.0
	MCV	fL	85.7	±5.0	91.5	±5.0	100.3	±6.0
	MCH	pg	26.4	±2.5	28.9	±2.5	31.3	±2.5
	MCHC	g/L	310	±30	318	±30	314	±30
	RDW-CV	%	18.3	±3.0	16.9	±3.0	16.0	±3.0
	RDW-SD	fL	54.7	±10.0	53.9	±10.0	54.9	±12.0
	PLT	×10 ⁹ /L	52	±20	250	±40	509	±60
	MPV	fL	10.6	±3.0	10.8	±3.0	9.7	±3.0
	PDW	fL	11.9	±3.0	14.3	±3.0	11.9	±3.0
	PCT	%	0.057	±0.050	0.275	±0.100	0.505	±0.200
P-LCR	%	27.8	±8.0	32.5	±8.0	24.8	±8.0	
P-LCC	×10 ⁹ /L	15	±15	83	±25	129	±35	
PDW	/	10.3	±3.0	10.3	±3.0	10.2	±3.0	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (Technical File Version A6.0 to A 6.5)	WBC	×10 ⁹ /L	3.33	±0.50	7.97	±1.00	17.55	±2.50
	Neu%	%	48.4	±9.0	57.1	±8.0	66.6	±7.0
	Lym%	%	39.5	±9.0	30.6	±8.0	20.8	±6.0
	Mon%	%	7.1	±4.0	7.3	±5.0	7.6	±6.0
	Eos%	%	5.0	±5.0	5.0	±5.0	5.0	±5.0
	Bas%	%	64.0	±8.0	72.7	±8.0	82.0	±8.0
	Neu#	×10 ⁹ /L	1.61	±0.40	4.55	±0.70	11.69	±1.40
	Lym#	×10 ⁹ /L	1.32	±0.40	2.44	±0.70	3.65	±1.10
	Mon#	×10 ⁹ /L	0.24	±0.14	0.58	±0.50	1.33	±1.10
	Eos#	×10 ⁹ /L	0.17	±0.15	0.40	±0.40	0.88	±0.88
	Bas#	×10 ⁹ /L	2.13	±0.30	5.79	±0.70	14.39	±1.50
	RBC	×10 ¹² /L	2.11	±0.18	4.44	±0.24	5.24	±0.50
	HGB	g/L	57	±4	130	±6	166	±8
	HCT	%	17.7	±2.0	39.6	±3.0	50.6	±4.0
	MCV	fL	84.2	±5.0	89.0	±5.0	96.5	±6.0
	MCH	pg	26.8	±2.5	29.3	±2.5	31.8	±2.5
	MCHC	g/L	317	±30	328	±30	328	±30
	RDW-CV	%	18.6	±3.0	17.2	±3.0	16.3	±3.0
	RDW-SD	fL	57.0	±10.0	55.7	±10.0	57.2	±12.0
	PLT	×10 ⁹ /L	49	±20	252	±40	536	±60
MPV	fL	10.3	±3.0	10.7	±3.0	9.5	±3.0	
PDW	fL	11.3	±3.0	13.9	±3.0	11.6	±3.0	
PCT	%	0.051	±0.050	0.271	±0.100	0.516	±0.200	
P-LCR	%	25.6	±8.0	31.6	±8.0	23.7	±8.0	
P-LCC	×10 ⁹ /L	15	±15	80	±25	128	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (Technical File Version A6.6 or higher)	WBC	×10 ⁹ /L	3.55	±0.50	8.35	±1.00	18.54	±2.50
	Neu%	%	49.2	±9.0	57.4	±8.0	66.6	±7.0
	Lym%	%	39.5	±9.0	30.5	±8.0	20.7	±6.0
	Mon%	%	7.3	±4.0	7.1	±5.0	7.7	±6.0
	Eos%	%	4.0	±4.0	5.0	±5.0	5.0	±5.0
	Bas%	%	64.6	±8.0	73.2	±8.0	82.5	±8.0
	Neu#	×10 ⁹ /L	1.75	±0.40	4.79	±0.70	12.35	±1.40
	Lym#	×10 ⁹ /L	1.40	±0.40	2.55	±0.70	3.84	±1.10
	Mon#	×10 ⁹ /L	0.26	±0.14	0.59	±0.50	1.43	±1.10
	Eos#	×10 ⁹ /L	0.14	±0.14	0.42	±0.42	0.93	±0.93
	Bas#	×10 ⁹ /L	2.30	±0.30	6.11	±0.70	15.29	±1.50
	RBC	×10 ¹² /L	2.17	±0.18	4.46	±0.24	5.36	±0.50
	HGB	g/L	58	±4	129	±6	167	±8
	HCT	%	17.9	±2.0	39.2	±3.0	50.9	±4.0
	MCV	fL	82.3	±5.0	87.8	±5.0	95.1	±6.0
	MCH	pg	27.0	±2.5	29.3	±2.5	31.7	±2.5
	MCHC	g/L	325	±30	332	±30	330	±30
	RDW-CV	%	18.0	±3.0	16.7	±3.0	15.7	±3.0
	RDW-SD	fL	54.6	±10.0	53.7	±10.0	54.6	±12.0
	PLT	×10 ⁹ /L	47	±20	247	±40	521	±60
MPV	fL	9.6	±3.0	9.9	±3.0	8.8	±3.0	
PDW	fL	9.9	±3.0	12.3	±3.0	10.0	±3.0	
PCT	%	0.046	±0.046	0.245	±0.100	0.458	±0.200	
P-LCR	%	21.5	±8.0	26.6	±8.0	19.0	±8.0	
P-LCC	×10 ⁹ /L	15	±15	66	±25	99	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D6-CRP DH76CRP (Technical File Version B1.0 to B2.0)	WBC	×10 ⁹ /L	3.44	±0.50	7.97	±1.00	17.78	±2.50
	Neu%	%	50.8	±9.0	58.3	±8.0	67.8	±7.0
	Lym%	%	39.0	±9.0	30.6	±8.0	19.8	±6.0
	Mon%	%	7.2	±4.0	7.1	±5.0	7.4	±6.0
	Eos%	%	3.0	±3.0	4.0	±4.0	5.0	±5.0
	Bas%	%	63.0	±8.0	71.9	±8.0	46.0	±8.0
	Neu#	×10 ⁹ /L	1.74	±0.40	4.65	±0.70	12.06	±1.40
	Lym#	×10 ⁹ /L	1.34	±0.40	2.44	±0.70	3.52	±1.10
	Mon#	×10 ⁹ /L	0.25	±0.14	0.57	±0.50	1.32	±1.10
	Eos#	×10 ⁹ /L	0.10	±0.10	0.32	±0.32	0.89	±0.89
	Bas#	×10 ⁹ /L	2.16	±0.30	5.73	±0.70	8.18	±1.50
	RBC	×10 ¹² /L	2.14	±0.18	4.42	±0.24	5.32	±0.50
	HGB	g/L	59	±4	128	±6	165	±8
	HCT	%	17.8	±2.0	38.8	±3.0	50.8	±4.0
	MCV	fL	83.0	±5.0	87.9	±5.0	95.4	±6.0
	MCH	pg	27.4	±2.5	29.0	±2.5	30.8	±2.5
	MCHC	g/L	329	±30	328	±30	321	±30
	RDW-CV	%	18.5	±3.0	17.1	±3.0	16.2	±3.0
	RDW-SD	fL	55.9	±10.0	54.4	±10.0	55.9	±12.0
	PLT	×10 ⁹ /L	51	±20	245	±40	518	±60
MPV	fL	9.0	±3.0	10.1	±3.0	8.9	±3.0	
PDW	fL	11.6	±3.0	12.8	±3.0	10.4	±3.0	
PCT	%	0.046	±0.046	0.245	±0.100	0.459	±0.200	
P-LCR	%	21.4	±8.0	27.7	±8.0	19.8	±8.0	
P-LCC	×10 ⁹ /L	15	±15	68	±25	102	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND D6-CRP DH76CRP (Technical File Version B2.2 or higher)	WBC	×10 ⁹ /L	3.53	±0.50	8.20	±1.00	17.81	±2.50
	Neu%	%	50.8	±9.0	57.8	±8.0	68.3	±7.0
	Lym%	%	37.7	±9.0	30.1	±8.0	19.7	±6.0
	Mon%	%	7.5	±4.0	7.1	±5.0	7.0	±6.0
	Eos%	%	4.0	±4.0	5.0	±5.0	5.0	±5.0
	Bas%	%	62.3	±8.0	72.2	±8.0	81.2	±8.0
	Neu#	×10 ⁹ /L	1.79	±0.40	4.74	±0.70	12.16	±1.40
	Lym#	×10 ⁹ /L	1.33	±0.40	2.47	±0.70	3.51	±1.10
	Mon#	×10 ⁹ /L	0.26	±0.14	0.58	±0.50	1.25	±1.10
	Eos#	×10 ⁹ /L	0.14	±0.14	0.41	±0.41	0.89	±0.89
	Bas#	×10 ⁹ /L	2.20	±0.30	5.92	±0.70	14.46	±1.50
	RBC	×10 ¹² /L	2.20	±0.18	4.53	±0.24	5.42	±0.50
	HGB	g/L	59	±4	129	±6	167	±8
	HCT	%	18.2	±2.0	39.7	±3.0	51.9	±4.0
	MCV	fL	82.9	±5.0	87.8	±5.0	95.7	±6.0
	MCH	pg	26.9	±2.5	28.6	±2.5	30.6	±2.5
	MCHC	g/L	323	±30	324	±30	319	±30
	RDW-CV	%	18.2	±3.0	16.8	±3.0	15.8	±3.0
	RDW-SD	fL	55.5	±10.0	54.4	±10.0	54.9	±12.0
	PLT	×10 ⁹ /L	52	±20	254	±40	533	±60
MPV	fL	9.4	±3.0	10.3	±3.0	9.2	±3.0	
PDW	fL	11.9	±3.0	13.3	±3.0	10.9	±3.0	
PCT	%	0.050	±0.050	0.262	±0.100	0.499	±0.200	
P-LCR	%	24.1	±8.0	29.7	±8.0	21.7	±8.0	
P-LCC	×10 ⁹ /L	15	±15	76	±25	117	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DF50 Vet DF52 Vet DF55 Vet DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	3.36	± 0.50	7.99	± 1.00	17.81	± 2.50
	Neu%	%	49.6	± 9.0	55.5	± 8.0	62.3	± 7.0
	Lym%	%	39.6	± 9.0	30.9	± 8.0	20.7	± 6.0
	Mon%	%	5.0	± 4.0	5.0	± 5.0	5.2	± 5.2
	Eos%	%	5.8	± 5.0	8.6	± 6.0	11.8	± 7.0
	Bas%	%	1.0	± 1.0	1.0	± 1.0	1.0	± 1.0
	Neu#	$\times 10^9/L$	1.67	± 0.40	4.43	± 0.70	11.10	± 1.40
	Lym#	$\times 10^9/L$	1.33	± 0.40	2.47	± 0.70	3.69	± 1.10
	Mon#	$\times 10^9/L$	0.17	± 0.14	0.40	± 0.40	0.93	± 0.93
	Eos#	$\times 10^9/L$	0.20	± 0.15	0.69	± 0.50	2.10	± 1.30
	Bas#	$\times 10^9/L$	0.03	± 0.03	0.08	± 0.08	0.18	± 0.18
	RBC	$\times 10^{12}/L$	2.26	± 0.18	4.59	± 0.24	5.48	± 0.50
	HGB	g/L	58	± 4	131	± 6	175	± 8
	HCT	%	19.0	± 2.0	41.3	± 3.0	54.0	± 4.0
	MCV	fL	84.1	± 5.0	89.9	± 5.0	98.5	± 6.0
	MCH	pg	25.7	± 2.5	28.6	± 2.5	31.9	± 2.5
	MCHC	g/L	308	± 30	320	± 30	325	± 30
	RDW-CV	%	14.1	± 3.0	13.4	± 3.0	12.7	± 3.0
	RDW-SD	fL	50.2	± 10.0	50.3	± 10.0	51.1	± 12.0
	PLT	$\times 10^9/L$	61	± 20	255	± 40	480	± 60
MPV	fL	9.1	± 3.0	9.7	± 3.0	8.7	± 3.0	
PDW	fL	10.8	± 3.0	14.2	± 3.0	11.9	± 3.0	
PCT	%	0.055	± 0.050	0.246	± 0.100	0.417	± 0.200	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (Technical File Version A4.0 to A5.7)	WBC	$\times 10^9/L$	3.30	± 0.50	7.88	± 1.00	17.78	± 2.50
	Neu%	%	49.8	± 9.0	55.9	± 8.0	62.9	± 7.0
	Lym%	%	39.1	± 9.0	30.5	± 8.0	20.1	± 6.0
	Mon%	%	5.2	± 4.0	5.0	± 5.0	5.2	± 5.2
	Eos%	%	5.9	± 5.0	8.6	± 6.0	11.8	± 7.0
	Bas%	%	1.5	± 1.5	1.5	± 1.5	2.0	± 2.0
	Neu#	$\times 10^9/L$	1.64	± 0.40	4.41	± 0.70	11.18	± 1.40
	Lym#	$\times 10^9/L$	1.29	± 0.40	2.40	± 0.70	3.57	± 1.10
	Mon#	$\times 10^9/L$	0.17	± 0.14	0.39	± 0.39	0.92	± 0.91
	Eos#	$\times 10^9/L$	0.19	± 0.15	0.68	± 0.50	2.10	± 1.30
	Bas#	$\times 10^9/L$	0.05	± 0.05	0.12	± 0.12	0.36	± 0.35
	RBC	$\times 10^{12}/L$	2.16	± 0.18	4.45	± 0.24	5.35	± 0.50
	HGB	g/L	58	± 4	129	± 6	169	± 8
	HCT	%	18.0	± 2.0	40.3	± 3.0	52.2	± 4.0
	MCV	fL	83.6	± 5.0	90.6	± 5.0	97.6	± 6.0
	MCH	pg	26.9	± 2.5	28.9	± 2.5	31.5	± 2.5
	MCHC	g/L	322	± 30	319	± 30	323	± 30
	RDW-CV	%	15.6	± 3.0	14.4	± 3.0	13.9	± 3.0
	RDW-SD	fL	54.3	± 10.0	53.8	± 10.0	55.1	± 12.0
	PLT	$\times 10^9/L$	53	± 20	256	± 40	472	± 60
MPV	fL	9.3	± 3.0	9.6	± 3.0	8.9	± 3.0	
PDW	fL	10.9	± 3.0	14.3	± 3.0	12.2	± 3.0	
PCT	%	0.053	± 0.050	0.247	± 0.100	0.445	± 0.200	
P-LCR	%	22.6	± 8.0	34.4	± 8.0	28.7	± 8.0	
P-LCC	$\times 10^9/L$	15	± 15	89	± 25	143	± 35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (Technical File Version A6.0)	WBC	×10 ⁹ /L	3.36	±0.50	8.16	±1.00	18.46	±2.50
	Neu%	%	49.6	±9.0	55.7	±8.0	62.3	±7.0
	Lym%	%	39.7	±9.0	30.4	±8.0	20.1	±6.0
	Mon%	%	5.1	±4.0	5.2	±5.0	5.7	±5.7
	Eos%	%	5.6	±5.0	8.7	±6.0	11.9	±7.0
	Bas%	%	1.5	±1.5	1.5	±1.5	1.5	±1.5
	Neu#	×10 ⁹ /L	1.67	±0.40	4.54	±0.70	11.50	±1.40
	Lym#	×10 ⁹ /L	1.34	±0.40	2.48	±0.70	3.71	±1.10
	Mon#	×10 ⁹ /L	0.17	±0.14	0.42	±0.42	1.05	±1.05
	Eos#	×10 ⁹ /L	0.19	±0.15	0.71	±0.50	2.20	±1.30
	Bas#	×10 ⁹ /L	0.05	±0.05	0.12	±0.12	0.28	±0.28
	RBC	×10 ¹² /L	2.15	±0.18	4.43	±0.24	5.26	±0.50
	HGB	g/L	58	±4	130	±6	170	±8
	HCT	%	17.8	±2.0	38.7	±3.0	49.6	±4.0
	MCV	fL	83.1	±5.0	87.4	±5.0	94.1	±6.0
	MCH	pg	27.0	±2.5	29.3	±2.5	32.2	±2.5
	MCHC	g/L	326	±30	336	±30	342	±30
	RDW-CV	%	15.8	±3.0	14.8	±3.0	14.2	±3.0
	RDW-SD	fL	55.7	±10.0	55.6	±10.0	57.4	±12.0
	PLT	×10 ⁹ /L	55	±20	249	±40	482	±60
MPV	fL	9.0	±3.0	9.6	±3.0	8.8	±3.0	
PDW	fL	10.3	±3.0	13.7	±3.0	11.7	±3.0	
PCT	%	0.052	±0.050	0.244	±0.100	0.437	±0.200	
P-LCR	%	26.0	±8.0	34.9	±8.0	27.9	±8.0	
P-LCC	×10 ⁹ /L	15	±15	89	±25	139	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low	Normal	High
DYMIND DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (Technical File Version A6.1 or higher)	WBC	×10 ⁹ /L	3.39 ±0.50	8.27 ±1.00	18.45 ±2.50
	Neu%	%	49.4 ±9.0	55.3 ±8.0	62.2 ±7.0
	Lym%	%	39.6 ±9.0	30.8 ±8.0	20.5 ±6.0
	Mon%	%	5.5 ±4.0	5.3 ±5.0	5.7 ±5.7
	Eos%	%	5.5 ±5.0	8.6 ±6.0	11.6 ±7.0
	Bas%	%	1.0 ±1.0	1.0 ±1.0	1.0 ±1.0
	Neu#	×10 ⁹ /L	1.68 ±0.40	4.57 ±0.70	11.47 ±1.40
	Lym#	×10 ⁹ /L	1.34 ±0.40	2.55 ±0.70	3.78 ±1.10
	Mon#	×10 ⁹ /L	0.19 ±0.14	0.44 ±0.44	1.05 ±1.05
	Eos#	×10 ⁹ /L	0.19 ±0.15	0.71 ±0.50	2.14 ±1.30
	Bas#	×10 ⁹ /L	0.03 ±0.03	0.08 ±0.08	0.18 ±0.18
	RBC	×10 ¹² /L	2.16 ±0.18	4.45 ±0.24	5.28 ±0.50
	HGB	g/L	58 ±4	129 ±6	170 ±8
	HCT	%	18.0 ±2.0	38.9 ±3.0	50.0 ±4.0
	MCV	fL	83.5 ±5.0	87.5 ±5.0	94.6 ±6.0
	MCH	pg	26.9 ±2.5	29.3 ±2.5	32.2 ±2.5
	MCHC	g/L	323 ±30	334 ±30	340 ±30
	RDW-CV	%	15.4 ±3.0	14.5 ±3.0	13.8 ±3.0
	RDW-SD	fL	54.5 ±10.0	54.2 ±10.0	55.4 ±12.0
	PLT	×10 ⁹ /L	46 ±20	249 ±40	530 ±60
	MPV	fL	9.3 ±3.0	9.6 ±3.0	8.7 ±3.0
PDW	fL	10.3 ±3.0	13.5 ±3.0	11.5 ±3.0	
PCT	%	0.043 ±0.043	0.240 ±0.100	0.461 ±0.200	
P-LCR	%	27.8 ±8.0	34.8 ±8.0	27.3 ±8.0	
P-LCC	×10 ⁹ /L	15 ±15	87 ±25	144 ±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND DM71X DM72X DM78X DM79X (Technical File Version A1.0 or higher)	WBC	×10 ⁹ /L	3.52	±0.50	8.19	±1.00	18.01	±2.50
	Neu%	%	51.3	±9.0	58.9	±8.0	66.1	±7.0
	Lym%	%	38.5	±9.0	29.7	±8.0	19.3	±6.0
	Mon%	%	7.2	±4.0	7.4	±5.0	7.6	±6.0
	Eos%	%	3.0	±3.0	4.0	±4.0	7.0	±7.0
	Bas%	%	63.1	±8.0	70.9	±8.0	81.3	±8.0
	Neu#	×10 ⁹ /L	1.81	±0.40	4.83	±0.70	11.90	±1.40
	Lym#	×10 ⁹ /L	1.36	±0.40	2.43	±0.70	3.48	±1.10
	Mon#	×10 ⁹ /L	0.25	±0.14	0.61	±0.50	1.37	±1.10
	Eos#	×10 ⁹ /L	0.11	±0.11	0.33	±0.33	1.26	±1.26
	Bas#	×10 ⁹ /L	2.22	±0.30	5.81	±0.70	14.64	±1.50
	RBC	×10 ¹² /L	2.18	±0.18	4.50	±0.24	5.42	±0.50
	HGB	g/L	60	±4	131	±6	169	±8
	HCT	%	18.0	±2.0	39.7	±3.0	52.1	±4.0
	MCV	fL	82.5	±5.0	88.1	±5.0	96.1	±6.0
	MCH	pg	27.5	±2.5	29.0	±2.5	31.0	±2.5
	MCHC	g/L	334	±30	328	±30	321	±30
	RDW-CV	%	18.1	±3.0	16.7	±3.0	15.7	±3.0
	RDW-SD	fL	54.8	±10.0	53.5	±10.0	54.3	±12.0
	PLT	×10 ⁹ /L	51	±20	248	±40	530	±60
MPV	fL	9.5	±3.0	10.5	±3.0	9.4	±3.0	
PDW	fL	12.0	±3.0	13.6	±3.0	11.2	±3.0	
PCT	%	0.049	±0.049	0.269	±0.100	0.506	±0.200	
P-LCR	%	24.7	±8.0	30.7	±8.0	22.6	±8.0	
P-LCC	×10 ⁹ /L	15	±15	79	±25	122	±35	

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High

DM-5D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

LOT BC0522

2022-04-11

2022-07-10

Applicable Instrument	Parameter	Unit	Low		Normal		High	
DYMIND UN71 Vet UN73 Vet (Technical File Version A1.0 or higher)	WBC	×10 ⁹ /L	3.43	±0.50	8.29	±1.00	18.12	±2.50
	Neu%	%	50.3	±9.0	57.3	±8.0	66.1	±7.0
	Lym%	%	38.1	±9.0	30.7	±8.0	21.5	±6.0
	Mon%	%	6.6	±4.0	7.0	±5.0	7.4	±6.0
	Eos%	%	4.0	±4.0	4.0	±4.0	4.0	±4.0
	Bas%	%	1.0	±1.0	1.0	±1.0	1.0	±1.0
	Neu#	×10 ⁹ /L	1.72	±0.40	4.75	±0.70	11.97	±1.40
	Lym#	×10 ⁹ /L	1.31	±0.40	2.55	±0.70	3.89	±1.10
	Mon#	×10 ⁹ /L	0.23	±0.14	0.58	±0.50	1.34	±1.10
	Eos#	×10 ⁹ /L	0.14	±0.14	0.33	±0.33	0.72	±0.72
	Bas#	×10 ⁹ /L	0.03	±0.03	0.08	±0.08	0.18	±0.18
	RBC	×10 ¹² /L	2.18	±0.18	4.50	±0.24	5.32	±0.50
	HGB	g/L	60	±4	129	±6	166	±8
	HCT	%	18.1	±2.0	40.1	±3.0	51.7	±4.0
	MCV	fL	83.0	±5.0	89.1	±5.0	97.2	±6.0
	MCH	pg	27.5	±2.5	28.9	±2.5	31.2	±2.5
	MCHC	g/L	331	±30	325	±30	321	±30
	RDW-CV	%	18.5	±3.0	17.1	±3.0	16.0	±3.0
	RDW-SD	fL	48.8	±10.0	47.7	±10.0	48.2	±12.0
	PLT	×10 ⁹ /L	44	±20	241	±40	479	±60
MPV	fL	9.6	±3.0	9.5	±3.0	8.6	±3.0	
PDW	fL	9.0	±3.0	11.6	±3.0	9.5	±3.0	
PCT	%	0.044	±0.044	0.241	±0.100	0.435	±0.200	
P-LCR	%	19.4	±8.0	24.5	±8.0	18.2	±8.0	
P-LCC	×10 ⁹ /L	15	±15	62	±25	92	±35	

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



Low



Normal



High