



DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0622

 2022-05-12

 2022-07-10

Applicable Instrument	Parameter	Unit	Reference Value
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A9.0 to A11.9)	WBC	$\times 10^9/L$	9.30
	RBC	$\times 10^{12}/L$	4.41
	HGB	g/L	132
	MCV	fL	89.1
	PLT	$\times 10^9/L$	254
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11)	WBC	$\times 10^9/L$	9.86
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	131
	MCV	fL	84.9
	PLT	$\times 10^9/L$	254
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.12 or higher)	WBC	$\times 10^9/L$	10.24
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	133
	MCV	fL	84.8
	PLT	$\times 10^9/L$	251
UN71, UN73, UN76, DH73 (Technical File Version A2.0 to A5.3)	WBC	$\times 10^9/L$	9.59
	RBC	$\times 10^{12}/L$	4.57
	HGB	g/L	133
	MCV	fL	89.6
	PLT	$\times 10^9/L$	252

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	NK0622
-----	--------

Mfg	2022-05-12
-----	------------

Exp	2022-07-10
-----	------------

Applicable Instrument	Parameter	Unit	Reference Value
UN71, UN73, UN76, DH73 (Technical File Version A5.5 to A6.2)	WBC	$\times 10^9/L$	9.32
	RBC	$\times 10^{12}/L$	4.42
	HGB	g/L	134
	MCV	fL	86.6
	PLT	$\times 10^9/L$	251
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.46
	RBC	$\times 10^{12}/L$	4.60
	HGB	g/L	134
	MCV	fL	86.3
	PLT	$\times 10^9/L$	253
DH73Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.57
	HGB	g/L	134
	MCV	fL	90.1
	PLT	$\times 10^9/L$	254
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A10.0 to A11.1) DF50 Vet, DF52 Vet (Technical File Version A2.0 to A7.0)	WBC	$\times 10^9/L$	9.07
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	132
	MCV	fL	87.0
	PLT	$\times 10^9/L$	263

【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0622

 2022-05-12

 2022-07-10

Applicable Instrument	Parameter	Unit	Reference Value
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A11.2 to A11.6)	WBC	$\times 10^9/L$	9.21
	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	133
	MCV	fL	88.4
	PLT	$\times 10^9/L$	255
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	9.19
	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	133
	MCV	fL	85.9
	PLT	$\times 10^9/L$	252
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.05
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	133
	MCV	fL	85.5
	PLT	$\times 10^9/L$	257
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.23
	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	133
	MCV	fL	90.8
	PLT	$\times 10^9/L$	257

【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0622

 2022-05-12

 2022-07-10

Applicable Instrument	Parameter	Unit	Reference Value
DH71, DH76 (Technical File Version A10.4)	WBC	$\times 10^9/L$	8.47
	RBC	$\times 10^{12}/L$	4.63
	HGB	g/L	133
	MCV	fL	86.8
	PLT	$\times 10^9/L$	257
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	9.48
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	134
	MCV	fL	84.9
	PLT	$\times 10^9/L$	247
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A3.0 to A5.11)	WBC	$\times 10^9/L$	9.32
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	133
	MCV	fL	89.3
	PLT	$\times 10^9/L$	254
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.19
	RBC	$\times 10^{12}/L$	4.46
	HGB	g/L	132
	MCV	fL	86.7
	PLT	$\times 10^9/L$	255

【NOTE】


1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0622

 2022-05-12

 2022-07-10

Applicable Instrument	Parameter	Unit	Reference Value
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	9.58
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	132
	MCV	fL	85.9
	PLT	$\times 10^9/L$	252
D6-CRP, DH76CRP (Technical File Version B1.0 to B2.0)	WBC	$\times 10^9/L$	9.20
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	132
	MCV	fL	85.6
	PLT	$\times 10^9/L$	251
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	9.25
	RBC	$\times 10^{12}/L$	4.56
	HGB	g/L	133
	MCV	fL	86.0
	PLT	$\times 10^9/L$	258
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	9.18
	RBC	$\times 10^{12}/L$	4.64
	HGB	g/L	134
	MCV	fL	88.8
	PLT	$\times 10^9/L$	257



【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **LOT** NK0622

 **2022-05-12**
 **2022-07-10**

Applicable Instrument	Parameter	Unit	Reference Value
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A4.0 to A5.7)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	132
	MCV	fL	88.1
	PLT	$\times 10^9/L$	250
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.0)	WBC	$\times 10^9/L$	9.39
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	133
	MCV	fL	85.4
	PLT	$\times 10^9/L$	255
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	9.60
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	132
	MCV	fL	85.7
	PLT	$\times 10^9/L$	254
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.33
	RBC	$\times 10^{12}/L$	4.56
	HGB	g/L	134
	MCV	fL	85.8
	PLT	$\times 10^9/L$	258


【NOTE】


1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **LOT** NK0622

 2022-05-12

 2022-07-10

Applicable Instrument	Parameter	Unit	Reference Value
UN71 Vet, UN73 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.48
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	132
	MCV	fL	87.7
	PLT	$\times 10^9/L$	247

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.