


	文件类型	技术文件	文件版本	V8.7.0		
	项目代号	/	页 码	第 0 页 共 7 页		
	文件编号	GL DM-56-02-en				
文件名称： <p style="text-align: center;">CALIBRATOR NK0722 2022-08-10</p> 适用范围：公司所有五分类仪器						
编制：		审核：		批准：		
日期：		日期：		日期：		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V8.7.0	/	新建		石丽岐	郑剑通	2022.06.24
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库					
	<input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0722

 2022-06-12

 2022-08-10

Applicable Instrument	Parameter	Unit	Reference Value
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A9.0 to A11.9)	WBC	$\times 10^9/L$	9.21
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	133
	MCV	fL	90.8
	PLT	$\times 10^9/L$	255
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11)	WBC	$\times 10^9/L$	9.84
	RBC	$\times 10^{12}/L$	4.46
	HGB	g/L	133
	MCV	fL	86.6
	PLT	$\times 10^9/L$	252
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.12 or higher)	WBC	$\times 10^9/L$	10.20
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	134
	MCV	fL	86.5
	PLT	$\times 10^9/L$	253
UN71, UN73, UN76, DH73 (Technical File Version A2.0 to A5.3)	WBC	$\times 10^9/L$	9.62
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	136
	MCV	fL	91.2
	PLT	$\times 10^9/L$	256



【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **LOT** NK0722

 **2022-06-12**
 **2022-08-10**

Applicable Instrument	Parameter	Unit	Reference Value
UN71, UN73, UN76, DH73 (Technical File Version A5.5 to A6.2)	WBC	$\times 10^9/L$	9.36
	RBC	$\times 10^{12}/L$	4.41
	HGB	g/L	135
	MCV	fL	87.8
	PLT	$\times 10^9/L$	248
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.57
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	136
	MCV	fL	87.2
	PLT	$\times 10^9/L$	252
DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.56
	HGB	g/L	135
	MCV	fL	91.3
	PLT	$\times 10^9/L$	251
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A10.0 to A11.1) DF50 Vet, DF52 Vet (Technical File Version A2.0 to A7.0)	WBC	$\times 10^9/L$	8.93
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	134
	MCV	fL	88.1
	PLT	$\times 10^9/L$	257

【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0722

 2022-06-12

 2022-08-10

Applicable Instrument	Parameter	Unit	Reference Value
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A11.2 to A11.6)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.62
	HGB	g/L	135
	MCV	fL	89.4
	PLT	$\times 10^9/L$	253
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	9.11
	RBC	$\times 10^{12}/L$	4.64
	HGB	g/L	134
	MCV	fL	86.9
	PLT	$\times 10^9/L$	253
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.22
	RBC	$\times 10^{12}/L$	4.63
	HGB	g/L	136
	MCV	fL	86.4
	PLT	$\times 10^9/L$	257
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.18
	RBC	$\times 10^{12}/L$	4.65
	HGB	g/L	135
	MCV	fL	92.4
	PLT	$\times 10^9/L$	256

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	NK0722
-----	--------

M	2022-06-12
---	------------

H	2022-08-10
---	------------

Applicable Instrument	Parameter	Unit	Reference Value
DH71, DH76 (Technical File Version A10.4)	WBC	$\times 10^9/L$	8.51
	RBC	$\times 10^{12}/L$	4.66
	HGB	g/L	135
	MCV	fL	87.7
	PLT	$\times 10^9/L$	257
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	9.39
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	136
	MCV	fL	86.2
	PLT	$\times 10^9/L$	249
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A3.0 to A5.11)	WBC	$\times 10^9/L$	9.14
	RBC	$\times 10^{12}/L$	4.54
	HGB	g/L	135
	MCV	fL	91.0
	PLT	$\times 10^9/L$	254
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.25
	RBC	$\times 10^{12}/L$	4.46
	HGB	g/L	134
	MCV	fL	88.1
	PLT	$\times 10^9/L$	260

【NOTE】


1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR

LOT NK0722

 2022-06-12

 2022-08-10

Applicable Instrument	Parameter	Unit	Reference Value
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	9.73
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	135
	MCV	fL	86.9
	PLT	$\times 10^9/L$	254
D6-CRP, DH76CRP (Technical File Version B1.0 to B2.0)	WBC	$\times 10^9/L$	9.25
	RBC	$\times 10^{12}/L$	4.48
	HGB	g/L	134
	MCV	fL	86.9
	PLT	$\times 10^9/L$	249
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	9.41
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	135
	MCV	fL	86.9
	PLT	$\times 10^9/L$	259
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	9.02
	RBC	$\times 10^{12}/L$	4.65
	HGB	g/L	136
	MCV	fL	90.2
	PLT	$\times 10^9/L$	258



【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **LOT** NK0722

 **2022-06-12**
 **2022-08-10**

Applicable Instrument	Parameter	Unit	Reference Value
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A4.0 to A5.7)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	134
	MCV	fL	89.6
	PLT	$\times 10^9/L$	253
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.0)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	135
	MCV	fL	86.3
	PLT	$\times 10^9/L$	254
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	9.54
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	134
	MCV	fL	87.0
	PLT	$\times 10^9/L$	256
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.25
	RBC	$\times 10^{12}/L$	4.57
	HGB	g/L	136
	MCV	fL	86.4
	PLT	$\times 10^9/L$	252

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT NK0722

2022-06-12

2022-08-10

Applicable Instrument	Parameter	Unit	Reference Value
UN71 Vet, UN73 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.51
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	134
	MCV	fL	88.3
	PLT	$\times 10^9/L$	247

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.